



Xenon light source



Microfluidic
Reagent Rotor



Ready-to-Use reagents.
No warm-up period
required.

7 minutes

Sample-to-answer in
3 simple steps and
approximately 7–10 minutes

18 month

18-month
shelf life

34 tests

Up to 34 tests
in a single run


Clinical Chemistry⁺

Chemistry Analyzer Pointcare® V3



MNCHIP

23 dedicated panel options covering 46 parameters



	Comprehensive Diagnostic Profile (25+9) Comprehensive Diagnostic Profile (23+7) Comprehensive Diagnostic Profile (20+7) Comprehensive Diagnostic Profile (19+5) Health Check Plus Profile (15+5) Health Check Profile (14+5) Health Check Profile (13+4) Preanesthetic Plus Profile (9+2) Preanesthetic Profile (7+3) Critical Care Profile (13+4) Liver Profile (11+4) Liver & Kidney Profile (9+4) Kidney Profile (11+6) Electrolyte Profile (4+2) Diabetes Profile (4) Feline Inflammation Profile (10+3) Canine Inflammation Profile (7+2) Equine Profile (15+4) Avian & Reptile (12+4) Large Animal (10+3) Coagulation Profile (4) Triple tests Profile (3) Ammonia Profile																							
TP	TP	TP	TP	TP	TP	TP	TP	TP	TP		TP	TP	TP			TP	TP	TP	TP	TP		TP		
ALB	ALB	ALB	ALB	ALB	ALB	ALB	ALB		ALB		ALB	ALB	ALB			ALB	ALB	ALB	ALB	ALB		ALB		
GLO*	GLO*	GLO*	GLO*	GLO*	GLO*	GLO*	GLO*		GLO*		GLO*	GLO*	GLO*			GLO*	GLO*	GLO*	GLO*	GLO*		GLO*		
ALB/GLO*	A/G*	A/G*	A/G*	A/G*	A/G*	A/G*	A/G*		A/G*		A/G*	A/G*	A/G*			A/G*	A/G*	A/G*	A/G*	A/G*		A/G*		
ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT	ALT										ALT		
AST	AST	AST	AST	AST		AST		AST			AST	AST						AST	AST	AST				
AST/ALT*	AAR*	AAR*	AAR*	AAR*		AAR*		AAR*			AAR*	AAR*												
GGT	GGT	GGT	GGT	GGT							GGT	GGT				GGT		GGT		GGT				
ALP	ALP	ALP	ALP	ALP	ALP	ALP	ALP	ALP	ALP		ALP					ALP		ALP		ALP		ALP		
TBIL	TBIL	TBIL	TBIL	TBIL	TBIL	TBIL	TBIL				TBIL	TBIL						TBIL				TBIL		
DBIL											DBIL													
IBIL*											IBIL*													
PT																					PT			
APTT																					APTT			
TT																					TT			
FIB																					FIB			
TBA	TBA	TBA	TBA	TBA	TBA						TBA					TBA			TBA					
CRE	CRE	CRE	CRE	CRE	CRE	CRE	CRE	CRE	CRE	CRE		CRE	CRE			CRE		CRE				CRE		
BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN	BUN			BUN		BUN		BUN		BUN		
BUN/CRE*	B/C*	B/C*	B/C*	B/C*	B/C*	B/C*	B/C*	B/C*	B/C*	B/C*		B/C*	B/C*			B/C*		B/C*				B/C*		
UA																			UA					
CK	CK	CK		CK		CK	CK	CK										CK	CK	CK		CK		
LDH								LDH										LDH						
AMY	AMY	AMY	AMY	AMY	AMY	AMY	AMY									AMY	AMY					AMY		
LPS	LPS	LPS	LPS													LPS	LPS							
GLU	GLU	GLU	GLU	GLU	GLU	GLU	GLU	GLU	GLU	GLU		GLU	GLU		GLU		GLU	GLU	GLU			GLU		
FRU	FRU														FRU									
TG	TG			TG											TG									
CHOL	CHOL	CHOL	CHOL	CHOL	CHOL	CHOL	CHOL				CHOL				CHOL							CHOL		
K+	K+	K+	K+		K+					K+			K+	K+				K+	K+					
Na+	Na+	Na+	Na+		Na+					Na+			Na+	Na+				Na+	Na+					
Na+/K+	Na+/K+	Na+/K+	Na+/K+		Na+/K+					Na+/K+			Na+/K+	Na+/K+				Na+/K+	Na+/K+					
Cl-	Cl-	Cl-	Cl-							Cl-			Cl-	Cl-					Cl-					
Ca	Ca	Ca	Ca	Ca	Ca	Ca	Ca			Ca			Ca			Ca	Ca	Ca	Ca	Ca		Ca		
P	P	P	P	P	P	P	P			P			P					P	P			P		
CaxP*	CaxP*	CaxP*	CaxP*	CaxP*	CaxP*	CaxP*	CaxP*			CaxP*			CaxP*					CaxP*	CaxP*			CaxP*		
Cl-/P*	Cl-/P*																							
Mg	Mg	Mg		Mg						Mg										Mg				
tCO ₂	tCO ₂	tCO ₂	tCO ₂	tCO ₂						tCO ₂			tCO ₂	tCO ₂				tCO ₂						
NH ₃																						NH ₃		
LAC	LAC	LAC								LAC														
pH										pH														
AG*	AG*	AG*	AG*							AG*			AG*	AG*										
fASS cCRP																fSAA	cCRP							
VLDL*	VLDL*																							

* Calculated test value # Coming soon